

The Twentieth Senior Technical Manager's Meeting  
of the Acid Deposition Monitoring Network in East Asia  
4-5 September 2019, Kuala Lumpur, Malaysia

## **Consideration of follow-up action on Inter-laboratory Comparison (ILC) project**

### **I. Introduction**

1. The Inter-laboratory Comparison (ILC) project was conducted among the analytical laboratories in participating countries of the Acid Deposition Monitoring Network in East Asia (EANET), based on the quality assurance/quality control (QA/QC) programs of EANET. The ILC project has implemented since 1998 for wet deposition, 1999 for soil, 2000 for inland aquatic environment and 2005 for dry deposition. The objectives of ILC are shown below. It covers the evaluation of analytical results, analytical equipment and its operating condition and other practices, Regarding the third item, the follow-up action after the ILC results are received is important.
  - (i) To recognize the analytical precision and accuracy of the measurement in each participating laboratory,
  - (ii) To give further opportunities to improve the quality of the analysis on wet deposition, dry deposition (filter pack method), soil and inland aquatic monitoring of EANET,
  - (iii) To improve reliability of analytical data through the assessment of suitable analytical methods and techniques.
2. For past 10 year data (2009-2018) of ILC project, the ratios of flagged data for wet deposition are relatively stable for both high (5-10%) and low (9-21%) concentration samples. The ratios of flagged data for dry deposition are also in certain range for both large (6-29%) and small (15-43%) quantity samples. The ratios of flagged data are relatively stable (9-16%) also for inland aquatic environment. While all of analytical data of 2 laboratories satisfied Data Quality Objectives (DQOs) for wet deposition, the percentages of the data within DQOs of 7 laboratories were below 70%. These situations imply that there are large variations of analytical uncertainty among EANET participating laboratories, and thus the additional follow up after the implementation of ILC for some laboratories would be necessary.

### **II. Current status of follow-up action and use of the ILC sample as working standard**

3. When site and laboratory audits for Japanese EANET sites are implemented, the ILC results for past 5 years are shown to local staffs and the reasons of outliers are discussed if necessary. Because the detection limit is one of importance factors for analytical performance, the past 5 year results of detection limits for ionic species are also discussed at laboratory audit. According to the discussions, key factors which cause large uncertainty and outlier for ion chromatography analysis were identified as follows; (i) Improper calibration curve, (ii) Contamination of the samples, apparatus, analyzer, auto sampler, (iii) Analytical Instability, (iv) unstable analytical condition. Improvement of these factor will lead to reduce analytical uncertainty. Moreover, preparation and understanding the

contents of Standard Operating Procedure (SOP) will avoid careless mistakes and keep constant analytical skills of laboratory and field operators.

4. The ILC sample in last year can be used for the working standard of ion species analysis to check analytical validity. It is recommended to prepare the table of prepared values and DQO ranges every year in order to check analytical validity of working standard. Example of the table of prepared values and DQO ranges is shown in Table 1 of Annex. If outlier data are found, it is necessary check each procedure and reanalyze the samples, if necessary. The results of working standard analysis which should be measured by each analytical run are reported to the national center of Japan. The reporting form is shown in Table 2 of Annex. These data can be checked long term stability and compatibility of ion analysis.

### **III. Draft time schedule in 2020 for follow-up action for EANET participating laboratories**

5. The follow-up action after the ILC results a for EANET participating laboratories is planned according to the below draft schedule. The effectiveness of the follow-up could be evaluated by reducing the ratios of flagged data.

<u>February 2020</u>	Submission deadline ILC survey in 2019
<u>April 2020</u>	Announcement of setting values and analytical results from the NC to participating laboratories of ILC
<u>May to June 2020</u>	Request to the lab which flagged data were reported for reanalysis and submission of wet deposition data
<u>July to August 2020</u>	Technical guidance by the NC in order to identify factors which cause large uncertainty and outlier for ion chromatography analysis
<u>September 2020</u>	Report of ILC survey in 2019 at STM21

Annex

Table 1 Prepared values of each parameter in artificial rainwater (171w and 172w)

The prepared values of each parameter in artificial rainwater of inter-laboratory comparison projects of EANET  
 ※ Those show the values of pH, EC and concentration of ions in sample No. 171w and No. 172w diluted by 100 times

each parameter / year	pH	EC [mS/m]	SO42- [μmol/l]	NO3- [μmol/l]	Cl- [μmol/l]	Na+ [μmol/l]	K+ [μmol/l]	Ca2+ [μmol/l]	Mg2+ [μmol/l]	NH4+ [μmol/l]
the diluted sample No. 171w	4.85	2.98	60.0	30.9	33.2	26.2	4.8	33.9	9.5	52.1
-15%	4.12	2.53	51.0	26.3	28.2	22.3	4.1	28.8	8.1	44.3
-10%	4.37	2.68	54.0	27.8	29.9	23.6	4.3	30.5	8.6	46.9
10%	5.34	3.28	66.0	34.0	36.5	28.8	5.3	37.3	10.5	57.3
15%	5.58	3.43	69.0	35.5	38.2	30.1	5.5	39.0	10.9	59.9
the diluted sample No. 172w	5.15	1.03	18.1	11.0	12.6	9.6	1.8	13.1	2.6	10.1
-15%	4.38	0.88	15.4	9.4	10.7	8.2	1.5	11.1	2.2	8.6
-10%	4.64	0.93	16.3	9.9	11.3	8.6	1.6	11.8	2.3	9.1
10%	5.67	1.13	19.9	12.1	13.9	10.6	2.0	14.4	2.9	11.1
15%	5.92	1.18	20.8	12.7	14.5	11.0	2.1	15.1	3.0	11.6

The prepared values of each parameter in artificial rainwater of inter-laboratory comparison projects of EANET  
 ※ Those show the values of pH, EC and concentration of ions in sample No. 171w and No. 172w diluted by 100 times

each parameter / year	pH	EC [μS/cm]	SO42- [mg/l]	NO3- [mg/l]	Cl- [mg/l]	Na+ [mg/l]	K+ [mg/l]	Ca2+ [mg/l]	Mg2+ [mg/l]	NH4+ [mg/l]
the diluted sample No. 171w	4.85	29.8	5.764	1.916	1.177	0.602	0.188	1.359	0.231	0.940
-15%	4.12	25.3	4.899	1.629	1.000	0.512	0.160	1.155	0.196	0.799
-10%	4.37	26.8	5.187	1.724	1.059	0.542	0.169	1.223	0.208	0.846
10%	5.34	32.8	6.340	2.108	1.295	0.663	0.206	1.495	0.254	1.034
15%	5.58	34.3	6.628	2.203	1.354	0.693	0.216	1.562	0.266	1.081
the diluted sample No. 172w	5.15	10.3	1.739	0.682	0.447	0.221	0.070	0.525	0.063	0.182
-15%	4.38	8.8	1.478	0.580	0.380	0.188	0.060	0.446	0.054	0.155
-10%	4.64	9.3	1.565	0.614	0.402	0.199	0.063	0.473	0.057	0.164
10%	5.67	11.3	1.913	0.750	0.491	0.243	0.077	0.578	0.070	0.200
15%	5.92	11.8	2.000	0.784	0.514	0.254	0.081	0.604	0.073	0.210

Table 2 Reporting form of working standard measurement  
(Working standard should be measured by each analytical run)

Item		Analytical method	Type of Instrument	Standard conc. (µmol/l)	Analysis results of working standard								
					Date	Setting (µmol/l)	Measured (µmol/l)	Date	Setting (µmol/l)	Measured (µmol/l)	Date	Setting (µmol/l)	Measured (µmol/l)
SO <sub>4</sub> <sup>2-</sup>	IC	IC	Integration	0.3, 0.6, 1.0, 3.1, 6.2	2018/6/29	44.5	44.6	2018/7/17	44.5	44.3	2018/8/17	44.5	44.1
					10.2	9.8	10.2	10.2	10.2	10.0			
NO <sub>3</sub> <sup>-</sup>	IC	IC	Integration	0.5, 1.0, 1.6, 4.8, 9.7	2018/6/29	21.0	21.2	2018/7/17	21.0	21.1	2018/8/17	21.0	20.5
					8.4	8.3	8.4	8.4	8.4	8.2			
Cl <sup>-</sup>	IC	IC	Integration	0.8, 1.7, 2.8, 8.5, 16.9	2018/6/29	32.3	32.6	2018/7/17	32.3	33.1	2018/8/17	32.3	31.1
					8.5	8.3	8.5	8.8	8.5	8.4			
NH <sub>4</sub> <sup>+</sup>	IC	IC	Integration	1.3, 3.3, 5.5, 16.6, 33.3	2018/6/29	31.5	31.5	2018/7/17	31.5	32.1	2018/8/17	31.5	31.2
					13.0	13.5	13.0	13.8	13.0	13.4			
Na <sup>+</sup>	IC	IC	Integration	1.3, 2.6, 4.3, 13.0, 26.1	2018/6/29	18.3	18.7	2018/7/17	18.3	18.5	2018/8/17	18.3	18.6
					6.5	6.6	6.5	6.5	6.5	6.5			
K <sup>+</sup>	IC	IC	Integration	0.8, 1.5, 2.6, 7.7, 15.3	2018/6/29	6.9	7.0	2018/7/17	6.9	6.8	2018/8/17	6.9	6.8
					1.7	1.8	1.7	1.5	1.7	1.6			
Ca <sup>2+</sup>	IC	IC	Integration	0.7, 1.5, 2.5, 7.5, 15.0	2018/6/29	28.8	28.0	2018/7/17	28.8	28.4	2018/8/17	28.8	28.6
					3.7	3.7	3.7	3.7	3.7	3.8			
Mg <sup>2+</sup>	IC	IC	Integration	1.2, 2.5, 4.1, 12.3, 24.7	2018/6/29	7.0	7.3	2018/7/17	7.0	6.8	2018/8/17	7.0	7.0
					1.8	2.0	1.8	1.7	1.8	1.8			
EC	EC	Conductivity Meter	HORIWA DS-15	0.01 mol/l	2018/6/29	2.39	2.37	2018/7/17	2.39	2.34	2018/8/17	2.39	2.40
					0.67	0.72	0.67	0.71	0.67	0.73			
pH	pH	pH/ION Meter	HORIBA F-54	pH7, pH4	2018/6/29	4.85	5.05	2018/7/17	4.85	5.07	2018/8/17	4.85	5.00
					5.30	5.54	5.30	5.47	5.30	5.38			

Unit of EC is mS/m

QC form of ion analysis  
Site : Bannu

Analysis date : 2018/6/29-2018/8/17

Sample ID : 11 ~ 17

Analysis lab : ACAP

Reporter:

Responsible person